

Quick Start Guide



YSI ProQuattro Water Quality Meter

This guide serves as a quick reference for operating a YSI ProQuattro. It is for your convenience & not intended to replace the full-length manual where additional information may be found.

**This unit was calibrated on the day we sent it to you.
(See Calibration Certificate)**

Setting-up your Pro Quattro

Step 1

Attach the Quattro Cable to the YSI meter

Step 2

Remove calibration cup & attach probe guard or flow cell over probes.

- If using probe guard, carefully & slowly lower the probe end of the cable to required height in the well.
- If using the flow cell, attach to pump & begin to pump sample through the flow cell. Please ensure that the Flow Cell is kept up right by using our Van Walt base plate located in the Peli case. Inaccurate readings will be caused if the flow cell is left horizontal.



Step 3

Press & hold Green button to turn on/off meter.

Step 4

Press: PROBE-> DISPLAY to show desired sensor reading on screen. Allow 5 - 10 minutes for DO probe to warm up.

Logging

Step 1

Press: FILE-> Data ID List -> ADD NEW to create personal site Select: Log interval → Enter Value

Step 2

Press: SYSTEM -> LOGGING-> CONTINUOUS MODE to begin automatic interval logging

Step 3

Press: ENTER to begin logging.

Step 4

Press: SITE in drop down menu to select your personal site.

Step 5

Press: ENTER to stop logging.

Step 6

Press: FILE-> VIEW DATA to view data.

Quick Start Guide



YSI ProQuattro Water Quality Meter

This guide serves as a quick reference for operating a YSI ProQuattro. It is for your convenience & not intended to replace the full-length manual where additional information may be found.

Calibrating your YSI ProQuattro

Step 1

Please contact Van Walt if you wish to calibrate your YSI.

Spot Checking

Step 1

We would recommend a quality spot check to ensure the best results. Used solutions will be charged for. Do not return any opened bottles of solution. Please call us for help or guidance.

Problems

Please contact Van Walt if you have any problems.



When not in use

Step 1

Please store your YSI in a warm dry place. Please do not keep the YSI in the peli-case if inside of the peli-case is very wet.

Step 2

When finished with unit, please rinse probes & fill the calibration cup to the mark (approx. 1 cm) with tap water (deionised water should never be used for storage!). Fit calibration cup by sliding it over the probes & screwing right until hand tight

Step 3

In order to charge the unit with the wall charger it must be switched on before plugging into the mains. If the battery has run out then it will not switch on. If this happens plug it in a usb port on a laptop or pc for 15 mins and then transfer to the wall charger once switched on, this will charge quicker