

## YSI556 Checklist and Spot Check

Date Received 27/6/11  
 YSI Number and Serial Number YSI 2 0512 2002 AD  
 Overall Condition Good

	Reading	Last Calibration Date
pH7 Buffer	7.20	14/6/11
pH4 Buffer	4.11	14/6/11
Conductivity 1413 µS/cm	1400	14/6/11
Redox 250mV Solution	233.7	14/6/11
DO (fresh air)	99.1	14/6/11
Temperature °C	23.77	-

This post rental check gives an indication of the status of the probes in the condition they were returned. For each parameter the probes are submerged in the applicable solution and the reading recorded. It is not a full calibration or proof that the probes are in good condition.

Notes: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

### Checklist

- YSI 556 meter
- 4 Meter/ 10 Meter Cable Assy
- Probe Guard
- Calibration Cup
- Manual CD
- Quick Start Guide
- Hard Case
- Flow Cell (including 2 o-rings)
- 2 x 122004 Flow Cell Nipples (sent w/flow cell)
- 2 x 122013 Flow Cell Nipples (sent w/flow cell)
- 4 x Spare 'C' Type Batteries
- Spare DO Membrane and Electrolyte
- Calibration solutions - if used mark:

### Comments

\_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Name: J HAWKINS

Signed: [Signature]

Date: 28/6/11